

SOHN'S DILATORS®

Catalog No. SOD-03

DESCRIPTION:

A set of SOHN'S DILATORS® includes (1) 30 mm, (1) 35 mm and (1) 40 mm dilators.

INDICATIONS:

SOHN'S DILATORS® are intended for the treatment of chronic anal fissures, severely painful acute fissures, or frequent fissure recurrences.

CONTRAINDICATIONS:

Incontinence or high grade anal stenosis.

INSTRUCTIONS FOR USE:

Treatment of anal fissure:

1. Perform under general, regional or local anesthesia. If using general anesthesia, use local also.
2. Consider prophylactic parental antibiotics for patients at risk for complications of bacteremia.
3. Perform a thorough rectal examination.
4. All dilators must be well lubricated when inserted.

For patients selected to receive the 40 mm anal dilator:

5. Insert the 30 mm dilator and hold in place for 1 minute.
6. Insert the 35 mm dilator and hold in place for 1 minute.
7. Insert the 40 mm dilator and hold in place for 6 minutes.

For other patients:

5. Gradually dilate the anus.
6. Insert the preferred maximum size dilator for 6 minutes.

WARNING:

Excessive force could result in damage to tissue and incontinence. Do not use the 40 mm dilator in women over 60 years old, in men over 65, in patients with anal stenosis that persists after regional or local anesthesia has been administered, or in any patient at high risk for incontinence.

CLEANING:

Do not use abrasive materials or brushes.

STERILIZATION:

Steam sterilize per ANSI/AAMI ST79:2006 "Comprehensive Guide to Steam Sterilization and Sterility Assurance in Health Care Facilities" using one of the following methods:

Method	Steam (Wrapped)	Steam (Wrapped)	Flash Steam (Unwrapped)
Cycle	Gravity	Prevacuum	Gravity
Temperature	270°F/132°C	270°F/132°C	270°F/132°C
Exposure	15 mins	5 mins	3 mins

SOHN'S DILATORS is a registered trademark of Tools for Surgery, LLC



We make it simple, by design.

Tools for Surgery, LLC
8 Technology Drive, #100 • East Setauket, New York 11733
PH (631) 444-4448 • FX (631) 689-2943

www.toolsforsurgery.com